



Street Address:

3356 Big Pine Trail, Ste C&D
Champaign, IL 61822 USA

Phone: +1-217-359-2344

E-Mail: CRM@aocs.org; **Web:** www.aocs.org

Certified Reference Materials

AOCS 0215-A

Report of the certification process for

MON 87751

Soybean Certified Reference Materials

First Batch

OECD Unique ID MON-87751-7

Denise Williams
Technical Services Manager

Tiffanie West
Technical Services Director



ISO 17034:2016
A2LA Certificate 3438.01

Legal Notice

Neither AOCS nor any person acting on behalf of AOCS is responsible for the use which might be made of the following information.

AOCS Mission Statement

AOCS advances the science and technology of oils, fats, proteins, surfactants, and related materials, enriching the lives of people everywhere.

More information regarding AOCS is available at <http://www.aocs.org>

Contents

Abstract	4
Acknowledgements	5
Glossary	6
Introduction.....	8
Material Processing.....	8
Trait Verification to Certify Presence of MON 87751	8
Certified Value and Measurement Uncertainty	10
Homogeneity	10
Stability.....	10
References	12

Abstract

This report describes the preparation and certification of the soybean CRM AOCS 0215-A produced by AOCS Technical Services in 2016. The CRMs have been prepared according to ISO 17034:2016 and are intended to serve as control material for third party testing of soybean for transformation events. Seed of MON 87751 was provided by Bayer CropScience, St. Louis, MO (hereinafter “Bayer CropScience”). The MON 87751 seed was milled by grinding the bulk source according to soybean processing protocols at Texas A&M University. The certified value of AOCS 0215-A was based on the purity of the bulk seed material and with 95% confidence, the true value is ≥ 978 g/kg. The powder was aliquoted and bottled in 27-mL glass headspace vials and sealed under a nitrogen gas environment at Illinois Crop Improvement Association. The presence of MON 87751 in the soybean was verified using event-specific, qualitative PCR analysis by Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory). CRM samples should be stored in a dry, sealed container at ambient or cooler conditions in the dark.

Acknowledgements

The authors would like to express sincere appreciation and gratitude to several individuals and their companies for support and guidance throughout this project. Thanks go to Jack Milligan, Bayer CropScience, for offering AOCS the opportunity to manufacture and distribute these products; to Richard Clough, Texas A&M University for processing the samples; to Sandra Harrison and Charlie Drennan at Illinois Crop Improvement Association for packaging the samples; and to Frank Spiegelhalter, Greg Ditta, E. Pearce Smith, and Daniel Thompson, Eurofins-GeneScan for event-specific, qualitative PCR analysis including the provision of information on running the analyses and interpreting the results.

Glossary

AOCS	American Oil Chemists' Society
Conventional Crop	Crop variety with no history of transgenic technology and is produced through traditional plant-breeding techniques that rely on selecting and mating parent plants possessing promising traits and repeatedly selecting for superior performance among their offspring
DNA	Deoxyribonucleic Acid is the linear, double-helix macromolecule that makes up the genetic material of most organisms
Detection Limit	Lowest level at which target DNA can be detected in a sample.
EC	European Commission
Genome	The full set of genes and associated DNA characteristic of an organism
ISO	International Organisation for Standardisation
GMO	Organism that has had genetic sequences modified using molecular-level techniques
PCR	Polymerase Chain Reaction: technique used to determine whether a sample of plant tissue contains a particular DNA sequence. PCR relies on primer sets that zero in on a particular target DNA sequence and a special DNA-copying enzyme (DNA polymerase) that makes enough copies of the target sequence for identification and measurement
Qualitative PCR	PCR methods that determine the presence or absence of a specific target DNA sequence at a particular level of detection

Quantitation Limit	Lowest level at which the amount of target DNA sequence in a sample can be reproducible.
Quantitative PCR	PCR methods that estimate the relative amount of target DNA sequence in a mixture of DNA molecules
Trait : MON 87751	Lepidopteran insect resistance

Introduction

Plant genetic modification is an extension of traditional plant breeding. It allows plant breeders to develop crops with specific traits including insect, disease, and herbicide resistance; processing advantages; and nutritional enhancement. An important component for identifying these new traits is a Certified Reference Material created from leaf, seed, or grain containing the new trait as well as a CRM created from the conventionally bred matrix. The European Commission has mandated that from 18 April 2004, a method for detecting a new event derived from transgenic technology and Certified Reference Material must be available before the EC will consider authorizing acceptance of a new crop derived from transgenic technology. Several nations outside Europe also require grain and ingredients to be labeled above a threshold level before accepting a shipment.

To meet the above regulatory requirements for GMO determination, AOCS 0215-A was manufactured from soybean according to ISO 17034:2016 and in accordance with EC No 1829/2003, EC No 641/2004 and EC No 619/2011. This CRM is available from AOCS.

Material Processing

MON 87751 soybean seeds used to prepare AOCS 0215-A were homozygous resulting from several cycles of self-pollination. Bayer CropScience delivered 20 kg of MON 87751 soybean seed to AOCS. The MON 87751 seed was milled by Texas A&M University according to industry standard soybean processing procedures. Illinois Crop Improvement Association was contracted for packaging the samples. The powder was aliquoted and bottled in 27-mL glass headspace vials and sealed under a nitrogen gas environment.

Trait Verification to Certify Presence of MON 87751

Prior to packaging, bulk seed powder samples were taken from randomly selected areas and depths to form a 3 kg composite sample in accordance with the International Seed Testing Association's (ISTA) Seed Science and Technology Rules for batches up to 500 kg, five (5) working samples of 10 g each were prepared from the composite sample and

sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025 Accredited laboratory) for event-specific, qualitative PCR analysis. The analyses performed by Eurofins-GeneScan, New Orleans, LA (an ISO 17025 Accredited laboratory) were used to verify the presence of MON 87751 (Table 1).

Table 1. Trait verification testing on random composite samples of MON 87751 soybean performed by Eurofins-GeneScan on bulk material provided by Bayer CropScience

Sample	MON 87751 Presence
Composite Sample 1	Positive
Composite Sample 2	Positive
Composite Sample 3	Positive
Composite Sample 4	Positive
Composite Sample 5	Positive

After the bulk material was packaged, the presence of the MON 87751 trait was assessed on five (5) random vials of AOCs 0215-A. AOCs used the Random Number Generator function of Microsoft Excel to select samples for verification of trait presence and to rule out degradation during packaging. AOCs 0215-A sample numbers 349, 380, 371, 76, and 5 were sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025 Accredited laboratory) for MON 87751 event-specific, qualitative PCR analysis (Table 2). This data confirms the presence of the MON 87751 in vials of AOCs 0215-A.

Table 2. Trait verification testing on AOCs 0215-A MON 87751 soybean performed by Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory).

Sample	MON 87751 Presence
AOCs 0215-A 349	Positive
AOCs 0215-A 380	Positive
AOCs 0215-A 371	Positive
AOCs 0215-A 76	Positive
AOCs 0215-A 5	Positive

Certified Value and Measurement Uncertainty

The genetic purity of the seed lot used to produce AOCS 0215-A was assessed by Bayer CropScience. A total of 717 soybean seeds were subjected to individual seed testing for the presence of MON 87751 by qualitative event-specific PCR. 717 of the 717 seeds tested positive for the presence of MON 87751.

Purity estimation was calculated using SeedCalc8 (Remund *et al.*, 2008) and the Certified Value corresponds to the lower bound of true % purity. The % purity in the sample was 100%, when 717 seeds were tested. Using a 95% confidence level, the true % purity of the MON 87751 seed lot was at least 97.8%. Consequently, with 95% confidence, the true value is ≥ 978 g/kg.

The Measurement Uncertainty was based on the lower bound of the true % purity. The measurement uncertainty is the expanded uncertainty with a coverage factor of 2 and confidence level of 95%. It is obtained by combining the uncertainties from the purity assessment ($u_{char,rel}$), the homogeneity assessment ($u_{bb,rel}$), the transport stability assessment ($u_{sts,rel}$) and the long-term stability assessment ($u_{lts,rel}$):

$$u_{CRM,rel} = \sqrt{u_{char,rel}^2 + u_{bb,rel}^2 + u_{sts,rel}^2 + u_{lts,rel}^2}$$
$$U_{CRM} = 2 \times u_{CRM,rel} \times 1000 \text{ g/kg}$$

The expanded measurement uncertainty for AOCS 0215-A is -22 g/kg.

Homogeneity

The homogeneity of AOCS 0215-A is related to the purity of the seeds. 717 out of 717 seeds tested positive for the MON 87751 soybean event. Based on the sample purity of 100%, as determined using SeedCalc8, the batch was considered to be homogeneous.

Stability

Time, temperature and light are regarded as the most relevant influences on the stability of CRM (Linsinger, *et al.*, 2001). The influence of light is mitigated by shipping and storing

the vials in boxes, thus minimizing the possibility of degradation due to light. The influence of temperature is mitigated by storing the vials in a temperature-controlled room, and shipping vials at ambient temperature.

The effect of temperature and time are investigated.

A transport (short-term) stability study is conducted to assess the stability of maize CRM during transport. The temperature and time conditions in the study cover the typical conditions and the not so rare situations. The outcome of the study is considered transferable to other CRMs of similar property. Samples were subject to 3 different temperatures (4 °C (fridge), 25 °C (ambient), 60 °C (oven)) for 4 different durations (0, 1, 2, and 4 weeks). The study concluded that samples are stable at 4 °C (fridge) and 25 °C (ambient) for 4 weeks. The estimated uncertainty contribution from transport (short-term) stability is 1.0%.

A long-term stability study is conducted to assess the stability of maize CRM during storage. Samples are stored at 25 °C (ambient) and the stability of the sample is monitored as long as the samples is available. The storage temperate studied is 25 °C and the length of time to be studied is 10 years. The outcome of the study is considered transferable to other CRMs of similar property. In the initial 1-year stability study, samples were subject the storage condition for 4 different durations (0, 1, 3, 6 and 12 months). The study concluded that samples are stable at 25 °C (ambient) for 12 months. The estimated uncertainty contribution from long-term stability is 0.42%.

Stability of these CRMs has been listed as 1 year from the certification date. The materials were processed and are stored at ambient temperature, under nitrogen gas, in 27-mL glass headspace vials. These materials are expected to be stable for longer than the estimated expiration date. The stability of the powder material will be reevaluated at time of expiration. If the samples still test positive for the presence of the intended trait, the certificates will be extended.

References

Eurofins-GeneScan; 2219 Lakeshore Drive, Suite 400, New Orleans, LA 70122;
Telephone: +1 504 297 4330 Toll Free: +1 866 535 2730 Fax: +1 504 297 4335
<https://www.eurofinsus.com/food-testing/testing-services/gmo/>

Illinois Crop Improvement Association, 3105 Research Road, Champaign, IL 61826;
Telephone: +1 217 359 4053 Fax: +1 217 359 4075; <http://www.ilcrop.com/index.htm>

ISO 17025:2005 and ISO 17025:2017, General Requirements for the Competence of
Testing and Calibration Laboratories

ISO 17034:2016 (E) General requirements for the competence of reference material
producers

International Seed Testing Association, International Rules of Seed Testing: Seed
Science and Technology Rules, 2012

Regulation (EC) No 1829/2003 of the European Parliament and of the Council of 22
September 2003 on genetically modified food and feed; <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX%3A32003R1829&from=en>

Remund K, Simpson R, Laffont J-L, Wright D, and Gregoire S. Seedcalc8. 2008.
<https://www.seedtest.org/en/statistical-tools-for-seed-testing-content---1--3449--1102.html>

Texas A&M University; Food Protein Research and Development Center; 373 Olsen Blvd;
College Station, TX 77845, USA; Telephone: +1 979 862 2262 Fax: +1 979 845 2744;
<http://foodprotein.tamu.edu/>