

Certificate of Analysis AOCS 0216-A, MON 87403 Maize OECD Unique ID MON-87403-1			
Certified Presence	Certified Value	Measurement Uncertainty	Test Method
MON 87403 maize present	≥ 991 g/kg	+9 g/kg, -22 g/kg	event-specific real-time PCR
<p>Description: This is the first batch of MON 87403 maize CRM prepared by AOCS for Monsanto Company. It was produced in February 2016. AOCS 0216-A was drawn from two seed lots whose sample purity values were 99.7% and 99.9%. The certified value is based on the seed lot with the lower sample purity of 99.7% (717 out of 719 seeds tested positive for MON 87403). With 95% confidence, the true value is ≥ 991 g/kg. The measurement uncertainty is truncated on the right side such that the value does not exceed 1000 g/kg. The measurement uncertainty is the expanded uncertainty with a coverage factor of 2.0 and confidence level of 95%. This material is for limited purposes only: see "Intended Use" and "Terms and Conditions."</p>			
<p align="center">This certificate is valid through: July 2022</p> <p>This validity may be extended when further evidence of stability becomes available. Customers will be notified by AOCS if a stability issue arises at AOCS Headquarters.</p> <p>Introduced: 19 April 2018. Revised: 26 April 2019, 24 July 2019, 15 October 2019, 11 May 2020, 23 June 2020, 21 June 2021, 9 July 2021, 5 August 2021</p>			
<p align="center">Technical Services Manager Denise Williams</p>		<p align="center">Chief Executive Officer Patrick J. Donnelly</p>	
<p>AOCS Mission: AOCS advances the science and technology of oils, fats, proteins, surfactants and related materials, enriching the lives of people everywhere.</p>			

Characterization

Product Description AOCS 0216-A has been prepared by AOCS from maize seed. AOCS 0216-A is available in 27-mL glass headspace vials containing approximately 10 g of material. Users are informed that this reference material has been produced from processed seed of MON 87403 maize delivered by Monsanto Company. MON 87403 maize seeds used to prepare AOCS 0216-A were hemizygous through successive breeding generations, and the donor for the MON 87403 maize event was the female parent.

Homogeneity The homogeneity of AOCS 0216-A is related to the purity of the seeds. 717 out of 719 seeds tested positive for the MON 87403 maize event. Based on the sample purity of 99.7%, as determined using SeedCalc8, the batch was considered to be homogeneous. This information was provided by Monsanto Company.

To further confirm homogeneity, ten vials of AOCS 0216-A was evaluated to determine the within-unit and between-unit property value. The property value is defined as the ratio between copies of the event specific target (MON 87403) and copies of a taxon-specific reference target (*hmg*). Variability of this ratio within CRM vials and between CRM vials is reported as the within-unit relative standard deviation (RSD_w), and the between-unit relative standard deviation (RSD_b). The results summarized below confirms the homogeneity of AOCS 0216-A.

CRM	RSD_w [%]	RSD_b [%]	$u^*_{bu,rel}$ [%]
AOCS 0216-A	2.0	n.c. ¹	1.1

¹n.c.: RSD_b cannot be calculated as $MS_{between} < MS_{within}$. In this situation, maximum hidden inhomogeneity ($u^*_{bu,rel}$) is provided as an alternative

In addition, ten packaged samples, 10 g each, were tested using event-specific qualitative PCR methods developed and validated by Monsanto Company. Test results received from Eurofins-GeneScan, New Orleans, LA (an ISO 17025 Accredited laboratory) for qualitative, event-specific analysis were all consistent with the reported presence of the MON 87403 trait.

Stability AOCS Certified Reference Materials are tested for stability on an annual or biennial basis, depending on the type of homogeneity testing that is required, and certificates may be extended based on the outcome of this testing. Customers may request extended certificates, but they are informed that results are based on samples that are obtained from AOCS' inventory. AOCS cannot guarantee the integrity of samples outside of our control.

Analytical Method Used for Certification



The Certified value is based on the purity of the MON 87403 material used in production of this CRM. Eurofins-GeneScan, New Orleans, LA (an ISO 17025 Accredited laboratory) performed qualitative event-specific real-time PCR for MON 87403 to validate the presence of the MON 87403 trait and quantitative event-specific real-time PCR for MON 87403 to determine the copy number ratio between the event-specific targets and the taxon-specific targets for the assessment of homogeneity. Purity and homogeneity results were used to determine the expanded measurement uncertainty of the certified value.

Warnings and Precautions

This product is for laboratory use only and is not for consumption. The user of this CRM should follow safety requirements and rules issued by voluntary organizations and government agencies expert in the field of laboratory safety.

Intended Use

This CRM, AOCS 0216-A, is intended for use as quality control material or calibrant in methods for the detection, identification, and/or quantification of biotechnology-derived events.

Instructions for Use

Upon receipt the product should be stored in a sealed container in the dark and at ambient or cooler conditions. The product may have settled during shipment, therefore, thoroughly mix the CRM before use to ensure homogeneity.

If the user of this CRM intends to use it multiple times, proper protocols must be followed to ensure that the sample retains its integrity. Use a clean laboratory spatula to remove the intended sample amount. After the sample has been removed, flush the headsapce of the vial with nitrogen gas, then replace the rubber stopper. Place a new 20 mm tear-off aluminum unlined seal on top of the rubber stop and crimp it to the vial by using a crimping tool. Store the CRM in the dark and at ambient or cooler conditions and repeat this process for all subsequent uses.

Sample size:

The recommended **sample intake** suitable for DNA extraction and real time PCR is 1 g.

Protocols have been followed to ensure that this CRM has the presence of the MON 87403 trait. Please follow all instructions on this certificate to prevent contamination and be sure to store the CRM under the proper conditions.



Note: The AOCS 0216-A certification report is available online and a paper copy will be supplied upon request.

Terms and Conditions (i– ix)

- i. The CRM AOCS 0216-A shall be used solely 1) in assays for (a) detecting the presence of MON 87403, or (b) quantification of MON 87403; or 2) for determining whether an assay cross-reacts with CRM AOCS 0216-A. CRM AOCS 0216-A shall be used for no other purpose. Specifically, the CRM may not be used to develop a detection method for MON 87403 nor trait(s) present therein. No other rights are conveyed by the sale of the CRM 0216-A to any purchaser, including any rights to any pending or granted Monsanto Company Patents or other Monsanto Company Intellectual Property that may protect the CRM or MON 87403 or trait(s) present therein or a detection method for MON 87403.
- ii. Neither the CRM AOCS 0216-A, nor any extract or portion thereof shall be resold or redistributed by any purchaser, unless the resale or redistribution is required by national law in force in the purchaser's country.
- iii. Neither the CRM AOCS 0216-A, nor any extract or portion thereof, shall be used for human or animal consumption or human or animal trials.
- iv. Neither the CRM AOCS 0216-A, nor MON 87403 DNA, nor any part of either of these, shall be used for transformation or breeding.
- v. No characterization or derivation of MON 87403, of the trait or traits present in CRM AOCS 0216-A, or of the CRM AOCS 0216-A shall be performed, except as allowed for in section (i).
- vi. All assay activities undertaken using the CRM AOCS 0216-A shall be conducted in strict compliance with all Applicable Laws governing such activities and shall comply with conditions of all permits and authorizations which may be required for such activities; and such activities shall be strictly limited to assays in contained facilities, for example, laboratories.
- vii. Prior to disposal of any used or excess CRM AOCS 0216-A or residues thereof, such material or residue must be treated in a manner that degrades MON 87403 in the CRM material, such as by autoclaving.
- viii. CRM AOCS 0216-A shall not be exported nor re-exported in violation of any Applicable Laws or without securing any necessary export or import clearances or permits.



- ix. The CRM AOCS 0216-A is provided for the purpose of identifying, detecting and quantifying the event and for no other purpose. AOCS has tested CRM AOCS 0216-A within the past 24 months and states that it has determined it to be of sufficient quality and fit for the purposes stated herein. No Monsanto Company warranty is provided, whether express or implied, in relation to the CRM AOCS 0216-A and Monsanto Company materials. Monsanto Company makes no representation or warranty that the use of the CRM AOCS 0216-A, whether before, or after the effective date of the applicable certificate, will not infringe any patent or other proprietary rights.

INTERNET COPY

