

Mail Address: AOCS, P.O. Box 17190, Urbana, IL 61803-7190 USA Street Address: AOCS, 2710 S. Boulder Drive, Urbana, IL 618026996 USA Phone: +1-217-359-2344; Fax: +1-217-351-8091; E-Mail: CRM@aocs.org; Web: www.aocs.org

Certified Reference Materials

AOCS 0306-B5

Report of the certification process for

Non-Modified

Canola Certified Reference Materials

Fifth Batch

Denise Williams Technical Services Manager Scott Bloomer Technical Director



Legal Notice

Neither AOCS nor any person acting on behalf of AOCS is responsible for the use which might be made of the following information.

AOCS Mission Statement:

AOCS advances the science and technology of oils, fats, proteins, surfactants, and related materials, enriching the lives of people everywhere.

More information regarding AOCS is available at http://www.aocs.org

Report of Certification for 0306-B5 Page 2 of 12 ©AOCS, 2020

Table of Contents

| Abstract | |
|------------------------------|----|
| Acknowledgements | 5 |
| Glossary | |
| Introduction | |
| Materials and Methods | 8 |
| Stability | 9 |
| Results and Discussion | |
| Sample Homogeneity | 10 |
| Prepared Sample Verification | 11 |
| References | 12 |

Abstract

This report describes the preparation and certification of the canola CRM AOCS 0306-B5 produced by AOCS Technical Services in 2016. The CRMs have been prepared according to ISO 17034:2016 and are intended to serve as control material for third party testing of canola for transformation events and for no other purpose. The purity of the Non-Modified canola was verified using event-specific, qualitative PCR analysis by Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory). AOCS 0306-B5 is available in 0.5 ml skirted screw-cap self-sealing tubes. The canola (Non-Modified; "Drakkar") DNA was extracted from clean leaves provided by BASF Agricultural Solutions Seed US LLC. The leaf tissue genomic DNA sample shall be stored in the self-sealing tube at +4 °C in the dark.

Acknowledgements

The authors would like to express sincere appreciation and gratitude to several individuals and their companies for support and guidance throughout this project. Thanks go to Ray Shillito and Benoit Maes, BASF Agricultural Solutions Seed US LLC, for offering AOCS the opportunity to manufacture and distribute these products; to Heather Waxdahl, SGS-Midwest for packaging the samples; and to Frank Spiegelhalter, Greg Ditta, E. Pearce Smith, and Daniel Thompson, Eurofins-GeneScan for event-specific, qualitative PCR analysis including the provision of information on running the analyses and interpreting the results.

Glossary

| AOCS | American Oil Chemists' Society | |
|-------------------|---|--|
| Conventional Crop | A related organism/variety, its components and/or products for which there is experience of establishing safety based on common use as food | |
| DNA | Deoxyribonucleic Acid is the linear, double-helix macromolecule that makes up the genetic material of most organisms | |
| Detection Limit | Lowest level at which target DNA can be detected in a sample. | |
| EC | European Commission | |
| Genome | The full set of genes and associated DNA characteristic of an organism | |
| GMO | Genetically modified/engineered organism: an organism in which the genetic material has been changed through modern biotechnology in a way that does not occur naturally by multiplication and/or natural recombination. | |
| ISO | International Organisation for Standardisation | |
| PCR | Polymerase Chain Reaction: technique used to determine whether a sample of plant tissue contains a particular DNA sequence. PCR relies on primer sets that bind to a particular target DNA sequence and a special DNA-copying enzyme | |
| | Report of Certification for 0306-B5 Page 6 of 12 | |

eport of Certification for 0306-B Page 6 of 12 ©AOCS, 2020

| | (DNA polymerase) that exponentially amplifies the target sequence for identification and measurement |
|--------------------|---|
| Qualitative PCR | PCR methods that determine the presence or absence of a specific target DNA sequence at a particular level of detection |
| Quantitation Limit | Lowest level at which the amount of target DNA sequence in a sample can be reliably quantitated |
| Quantitative PCR | PCR methods that estimate the relative amount of target DNA sequence in a mixture of DNA molecules |

Report of Certification for 0306-B5 Page 7 of 12 ©AOCS, 2020

Introduction

Plant biotechnology is an extension of traditional plant breeding. It allows plant breeders to develop crops with specific traits including insect, disease, and herbicide resistance; processing advantages; and nutritional enhancement. An important component for identifying these new traits is a Certified Reference Material created from leaf, seed, or grain containing the new trait as well as a CRM created from the conventionally bred matrix. The European Commission has mandated that from 18 April 2004, a method for detecting a new event derived from modern biotechnology and a CRM must be available before the EC will consider authorizing acceptance of a new crop derived from modern biotechnology. Several nations outside Europe also require grain and ingredients to be labeled above a threshold level ranging from 0.90 to 5% of authorized biotech events before accepting a shipment.

To meet the above regulatory requirements for measurement standards, AOCS 0306-B5 was manufactured according to ISO 17034:2016 and in accordance with EC No 1829/2003. The CRMs are available from AOCS.

Materials and Methods

BASF Agricultural Solutions Seed US LLC delivered 2.0 mg of Non-modified canola ("Drakkar") DNA to AOCS. Five (5) working samples of DNA (10 μ g) each were prepared from the composite and sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory) for event-specific, qualitative PCR analysis to screen for the presence of the intended event, Non-Modified. This testing was for purity as well as homogeneity purposes.

The source leaf material was taken from plants which had been tested individually using a number of quality standards and was grown from seeds harvested from plants that had themselves passed the same criteria. Plants not meeting the quality standards were removed and destroyed. Leaf material was harvested from the plants which met the quality standards and frozen immediately, stored at -70 °C.

Report of Certification for 0306-B5 Page 8 of 12 ©AOCS, 2020 The genomic DNA was extracted from leaves of one or more plants according to CTABbased (Doyle JJ and Doyle JL, 1987) protocol. The integrity and concentration of the genomic DNA was determined by electrophoresis in a 1.0% agarose gel and ethidium bromide-staining and compared to lambda molecular weight standards by digital imaging quantification. The concentration measurement was done in triplicate, repeated in three different gels. No indications for physical degradation were apparent and the DNA migrated at positions higher than 40 Kb.

The leaf used to manufacture the non-modified materials was shown to be absent of Ms1, Ms8, Ms11, Rf1, Rf2, Rf3, T45, Topas 19/2 using PCR protocols at BASF Agricultural Solutions Seed US LLC. The non-modified canola leaf DNA was packaged by SGS-Midwest Seed Services in sterile, 0.5 ml skirted screw-cap self-sealing tubes in aliquots of 10 μ g.

AOCS used the Random Number Generator function of Microsoft Excel to select samples for verification of purity, and homogeneity. Sample numbers AOCS 0306-B5: 15, 26, 43, 95, and 135 were sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory) for event-specific, qualitative PCR analysis to screen for Non-Modified presence in the samples.

Stability

Stability of these CRMs has been listed as 1 year from the certification date. The materials were sealed and stored under refrigerated conditions, therefore not exposed to air and are expected to be stable for longer than the estimated expiration date. The stability of the leaf tissue genomic DNA material will be reevaluated annually. If the samples are still representative of the certified value, the certificates will be extended.

Report of Certification for 0306-B5 Page 9 of 12 ©AOCS, 2020

Results and Discussion

Sample Homogeneity

The purity data for the non-modified homogeneity samples is presented in Table 1.

Table 1. Results of the homogeneity testing performed by Eurofins-GeneScan on the Non-Modified bulk material ("Drakkar") provided by BASF Agricultural Solutions Seed US LLC

| Sample | Ms1, Ms8, Ms 11, Rf1, Rf2, Rf3, T45, Topas 19/2 Presence |
|----------------------|---|
| Homogeneity Sample 1 | Negative |
| Homogeneity Sample 2 | Negative |
| Homogeneity Sample 3 | Negative |
| Homogeneity Sample 4 | Negative |
| Homogeneity Sample 5 | Negative |

Prepared Sample Verification

Once the bulk material was processed and packaged, five (5) samples were identified by the Microsoft Excel Random Number Generator and sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025 accredited laboratory) for event-specific, qualitative PCR analysis. These results are presented in Table 2. These data show no contamination occurred during the packaging of AOCS 0306-B5. These results are in agreement with the homogeneity data presented in Table 1.

| Table 2. Results for the verification of AOCS 0306-B5 [Non-Modified canola | | |
|--|--|--|
| ("Drakkar")] material as tested by Eurofins-GeneScan with Non-Modified | | |
| event-specific, qualitative PCR analysis. | | |
| | | |

| Sample | Ms1, Ms8, Ms11, Rf1, Rf2, Rf3, T45, Topas 19/2 Presence |
|------------------|---|
| AOCS 0306-B5 15 | Negative |
| AOCS 0306-B5 26 | Negative |
| AOCS 0306-B5 43 | Negative |
| AOCS 0306-B5 95 | Negative |
| AOCS 0306-B5 135 | Negative |

References

Center for Environmental Risk Assessment GM Database http://www.cera-gmc.org/?action=gm_crop_database

Eurofins-GeneScan; 2219 Lakeshore Drive, Suite 400, New Orleans, LA 70122; Telephone: +1 504 297 4330 Toll Free: +1 866 535 2730 Fax: +1 504 297 4335 http://www.gmotesting.com

International Seed Testing Association, International Rules of Seed Testing: Seed Science and Technology Rules, 2012

ISO 17025:2005 and ISO 17025:2017, General Requirements for the Competence of Testing and Calibration Laboratories

ISO Guide 174034:2016 (E) General Requirements for the Competence of Reference Material Producers

SGS-Midwest, 236 32nd Avenue, Brookings, South Dakota 57006 Telephone: + 1 605 692 7611 Toll Free: +1 877 692 7611 Fax: +1 605 692 7617 http://www.mwseed.com/

Report of Certification for 0306-B5 Page 12 of 12 ©AOCS, 2020