



Street Address:

AOCS, 2710 S. Boulder Drive
Urbana, IL 61802- 6996 USA

Phone: +1-217-359-2344; **Fax:** +1-217-351-8091;

E-Mail: CRM@aoacs.org; **Web:** www.aocs.org

Certified Reference Materials

AOCS 0411-C2

Report of the certification process for

Non-modified

Maize Certified Reference Materials

Second Batch

Denise Williams
Technical Services Manager

Scott Bloomer
Technical Director

Version: 1.1 | Published: 31 March 2020



ISO 17034:2016
A2LA Certificate 3438.01

Legal Notice

Neither AOCS nor any person acting on behalf of AOCS is responsible for the use which might be made of the following information.

AOCS Mission Statement

AOCS advances the science and technology of oils, fats, proteins, surfactants, and related materials, enriching the lives of people everywhere.

More information regarding AOCS is available at <http://www.aocs.org>

Table of Contents

Abstract.....	4
Acknowledgements.....	5
Glossary.....	6
Introduction	8
Material Processing and Particle Size Analyses	8
Certified Value Assignment.....	11
Homogeneity Testing.....	11
Trait Verification.....	13
Stability	14
References.....	16

Abstract

This report describes the preparation and certification of the maize Certified Reference Material (CRM) AOCS 0411-C2 produced by AOCS Technical Services in 2019. The CRMs have been prepared by AOCS according to ISO 17034:2016 and are intended to serve as control material for third party testing of maize for transformation events. Non-modified maize seed powder (Breeding line: NP2171/NP2460) was provided by Syngenta Crop Protection, LLC and was prepared by grinding the bulk source at AVEKA, Inc., Woodbury, MN (an ISO 9001:2015 accredited facility). The non-modified maize seed powder was then aliquoted and packaged under a nitrogen gas environment at Illinois Crop Improvement Association (an ISO 17025:2017 accredited facility). The certified mass value of non-modified CRM AOCS 0411-C2 was based on the absence of genetically modified impurities and with 95% confidence, the true value is lower than 1 g/kg. Homogeneity testing was performed at Eurofins-GeneScan, New Orleans, LA (an ISO 17025:2005 accredited laboratory) using quantitative real-time PCR after the CRM AOCS 0411-C2 was bottled. Homogeneity results indicated that CRM AOCS 0411-C2 is homogenous and were used to verify the absence of the Event 5307 in this CRM. CRM AOCS 0411-C2 is available in 27-mL glass headspace vials. This CRM shall be stored dry in a sealed container at ambient or cooler conditions in the dark.

Acknowledgements

The authors would like to express sincere appreciation and gratitude to several individuals and their companies for support and guidance throughout this project. Thanks go to Kristina Burgin, Syngenta Crop Protection, LLC, for offering AOCS the opportunity to manufacture and distribute these products; to Sandra Harrison and Charlie Drennan at Illinois Crop Improvement Association for packaging the samples; and to Frank Spiegelhalter, Greg Ditta, E. Pearce Smith, and Daniel Thompson, Eurofins-GeneScan for event-specific, real-time PCR analysis including the provision of information on running the analyses and interpreting the results.

Glossary

AOCS	American Oil Chemists' Society
Conventional Crop	Conventional counterpart means a related organism/variety, its components and/or products for which there is experience of establishing safety based on common use as food
Cycle threshold (Ct)	Number of PCR cycles required for the fluorescent signal to cross a threshold that exceeds background level
Detection Limit	Lowest level at which target DNA can be detected in a sample.
DNA	Deoxyribonucleic Acid is the linear, double-helix macromolecule that makes up the genetic material of most organisms
EC	European Commission
Genome	The full set of genes and associated DNA characteristic of an organism
GMO	Genetically modified/engineered organism: an organism in which the genetic material has been changed through modern biotechnology in a way that does not occur naturally by multiplication and/or natural recombination.
ISO	International Organisation for Standardisation
PCR	Polymerase Chain Reaction: technique used to determine whether a sample of plant tissue contains a particular DNA sequence. PCR relies on primer sets that bind to a particular target DNA sequence and a special DNA-copying enzyme (DNA polymerase) that exponentially amplifies the target sequence for identification and measurement

Qualitative PCR	PCR methods that determine the presence or absence of a specific target DNA sequence at a particular level of detection
Quantitation Limit	Lowest level at which the amount of target DNA sequence in a sample can be reliably quantitated
Quantitative PCR	PCR methods that estimate the relative amount of target DNA sequence in a mixture of DNA molecules
RSD _r	Relative standard deviation
SD	Standard deviation

Introduction

Plant biotechnology is an extension of traditional plant breeding. It allows plant breeders to develop crops with specific traits including insect, disease, and herbicide resistance; processing advantages; and nutritional enhancement. An important component for identifying these new traits is a Certified Reference Material (CRM) created from leaf, seed, or grain containing the new trait as well as a CRM created from the conventionally bred matrix. The European Commission has mandated that from 18 April 2004, a method for detecting a new event derived from transgenic technology and Certified Reference Material must be available before the EC will consider authorization of a new crop derived from transgenic technology. Several nations outside Europe also require grain and ingredients to be labeled above a threshold level before accepting a shipment.

To meet the above regulatory requirements for GMO determination, CRM AOCS 0411-C2 was manufactured from maize seed according to ISO 17034:2016 and in accordance with EC No 1829/2003. The CRM is available from AOCS.

Material Processing and Particle Size Analyses

The non-modified maize breeding line NP2171/NP2460 was used in the production of CRM AOCS 0411-C2. Seed from NP2171/NP2460 maize was first milled and analyzed for particle size distribution at AVEKA, Inc., Woodbury, Minnesota (an ISO 9001:2015 accredited facility).

Bulk seed received by AVEKA, Inc. from Syngenta Crop Protection, LLC was milled in a Fitzmill cryogenic hammermill using first a 690 μm screen. To further reduce particle size, this ground material was milled again under the same conditions using a 510 μm screen. The material was blended in a Patterson-Kelley V-blender, and after homogenization six samples taken at random were subject to particle size analyses using a Horiba LA-950 Light Scattering Particle Analyzer. For each sample, the particle size mean and range, and the percentage of particles below a certain size were calculated (Table 1). On average, the particle size of CRM AOCS 0411-C2 was $136.51 \pm 6.52 \mu\text{m}$, and 90% of the particles (i.e. D90) were smaller than $282.07 \pm 12.63 \mu\text{m}$.

**Table 1. Results of Particle Size Analyses of CRM AOCS 0411-C2
Conducted by AVEKA, Inc.**

	Sample 1 (μm)	Sample 2 (μm)	Sample 3 (μm)	Sample 4 (μm)	Sample 5 (μm)	Sample 6 (μm)	Average (μm)	Standard Deviation (μm)
Mean	126.02	135.90	145.55	140.62	134.67	136.32	136.51	6.52
Range	0.88-517	0.88-452	0.88-592	0.88-517	0.88-452	0.88-452	N/A ^(a)	N/A
D10 ^(b)	14.72	15.64	15.47	16.50	16.63	16.22	15.86	0.73
D50 ^(b)	109.41	122.63	129.91	128.29	122.08	124.63	122.83	7.27
D90 ^(b)	264.08	281.48	302.12	288.21	276.62	279.89	282.07	12.63

^(a) N/A = not applicable

^(b) D10, D50 and D90 indicate that 10%, 50% or 90% of the particles, respectively, are smaller than size given in table

The particle size distribution for each of the samples analyzed is presented as a histogram, with the x-axis showing discrete size bins up to 592.4 μm (Figure 1). Figure 1-A represents the percentage of particles of a given size, and Figure 1-B represents the cumulative particle size distribution, which reflects the total percentage of particles smaller than a given size. For all samples analyzed, 100% of particles were $\leq 592.4 \mu\text{m}$.





Figure 1. Particle size distribution plots. (A) Percentage of particles of a given size. (B) Cumulative distribution of particle sizes.

Bulk, non-modified maize seed powder for the production of CRM AOCS 0411-C2 was delivered to AOCS and it was then aliquoted and packaged in 27-ml glass headspace vials and sealed under a nitrogen gas environment at the Illinois Crop Improvement Association (an ISO 17025:2017 accredited facility).

Certified Value and Measurement Uncertainty Assignment

The genetic purity of the seed lot used to produce AOCS 0411-C2 was assessed by Syngenta Crop Protection, LLC. A total of 3000 maize seeds (10 aliquots of 300 seeds) were subjected to individual seed testing for the presence of 5307 maize by qualitative event-specific PCR. None of the NP2171/NP2460 3000 seeds tested positive for the presence of 5307 maize.

Purity estimation was calculated using SeedCalc8 (Remund *et al.*, 2008). The % impurity in the sample was 0% when 3000 seeds were tested. With 95% confidence, the true certified value is < 1 g/kg.

The measurement uncertainty (U_{CRM}) is the expanded uncertainty with a coverage factor of 1.65 and a confidence level of 95%. It is obtained by combining the uncertainties from the purity assessment ($u_{char,rel}$) and the homogeneity assessment ($u_{bb,rel}$):

$$u_{CRM,rel} = \sqrt{u_{char,rel}^2 + u_{bb,rel}^2}$$
$$U_{CRM} = 1.65 \times u_{CRM,rel} \times \text{purity estimation} \times 1000 \text{ g/kg}$$

The expanded measurement uncertainty for this CRM is .5 g/kg.

Homogeneity Testing

The material used for the production of CRM AOCS 0411-C2, non-modified maize, contains 0% impurities and is expected to be homogenous. After NP2171/NP2460 seed was ground and bottled as described above, ten samples of CRM AOCS 0411-C2 were randomly selected using the Microsoft Excel Random Number Generator function and

were sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025:2005 accredited laboratory) for homogeneity testing using quantitative real-time PCR.

Homogeneity was also assessed after bottling of AOCS 0411-C2 CRM using a 5307-specific, quantitative real-time PCR method (https://gmo-crl.jrc.ec.europa.eu/summaries/2014-12-05_VP_EURL-VL-07-11FINAL.pdf). A total of 10 samples of CRM AOCS 0411-C2 maize were analyzed, and for each sample, 2 independent DNA extractions and quantifications were performed at Eurofins-GeneScan using a test portion of 1 gram. Extracted DNA was checked for integrity by gel-electrophoresis and quantified prior to using it in quantitative real-time PCR. For each of the DNA extracts, all PCR reactions were done in triplicate.

The cycle threshold (Ct) values for an endogenous *adh1* maize gene and for event 5307 were used to calculate the number of copies (cp#) for either target. Subsequently, the ratio between event 5307 copy number and *adh1* copy number (5307 cp#/*adh1* cp#) was calculated and used to estimate the within-unit relative standard deviation (RSD_w) and between-unit relative standard deviation (RSD_b).

Within-unit relative standard deviation (RSD_w), between-unit relative standard deviation (RSD_b) were calculated as:

$$\text{Within-unit RSD: } RSD_w = \frac{\sqrt{MS_{within}}}{\bar{y}}$$

$$\text{Between-unit RSD: } RSD_b = \frac{\sqrt{\frac{MS_{between} - MS_{within}}{n}}}{\bar{y}}$$

where,

MS_{within}	within-unit mean square from an ANOVA
$MS_{between}$	between-unit mean square from an ANOVA
\bar{y}	mean of all results of the homogeneity study
n	mean number of replicates per unit

Table 2. The within-unit relative standard deviation (RSD_w), and the between-unit relative standard deviation (RSD_b) for vials of AOCS 0411-C2.

CRM	RSD _w [%]	RSD _b [%]
AOCS 0411-C2	Below LOQ	Below LOQ

The CRM will be determined to be homogeneous if the within-unit relative standard deviation (RSD_w) and between-unit relative standard deviation (RSD_b) are both ≤20%. No amplification was observed from non-modified CRM AOCS 0411-C2 DNA extracts when using the PCR method specific to Events 5307. Based on the quantitative real-time PCR analyses conducted, it was concluded that CRM AOCS 0411-C2 is homogenous (Table 2) because all results are below LOQ. These results are in agreement with homogeneity results from qualitative real-time PCR analyses and with the purity estimate for material non-modified maize calculated in the Certified Value and Measurement Uncertainty section above.

Trait Verification

The absence of Event 5307 in non-modified NP2171/NP2460 maize material was assessed in the same ten CRM AOCS 0411-C2 samples that were analyzed for homogeneity using 5307-specific, quantitative PCR analysis. Quantitative results were converted to qualitative data, and the results are presented in Table 3. In all instances Event 5307 was absent.

Table 3. Qualitative results for the homogeneity and verification of CRM AOCS 0411-C2 non-modified maize as tested by Eurofins-GeneScan with a 5307-specific, quantitative PCR analysis.	
Sample	Event 5307 Presence
AOCS 0411-C2 80	Negative
AOCS 0411-C2 108	Negative
AOCS 0411-C2 146	Negative
AOCS 0411-C2 199	Negative
AOCS 0411-C2 225	Negative

AOCS 0411-C2 317	Negative
AOCS 0411-C2 328	Negative
AOCS 0411-C2 350	Negative
AOCS 0411-C2 363	Negative
AOCS 0411-C2 398	Negative

Stability

Time, temperature and light are regarded as the most relevant influences on the stability of CRM (Linsinger, et al., 2001). The influence of light is mitigated by shipping and storing the vials in boxes, thus minimizing the possibility of degradation due to light. The influence of temperature is mitigated by storing the vials in a temperature controlled room, and shipping vials at ambient temperature. Therefore, only the influence of time need be investigated.

CRM stability over time will be analyzed by repeating the homogeneity study described above at a chosen shelf life of approximately every 24 months. The 24-month shelf life of CRM is chosen because the influence of analytical variation can be reduced by increasing the length of the stability study (Linsinger, et al., 2001).

The initial ratio between the number of copies of the GM event and the number of copies of the endogenous reference gene from the homogeneity study will establish the base line for the stability study. The ratio at each 24-month interval will be compared to the ratio established in the homogeneity study. The CRM will be determined to be stable if the variability of the ratios, determined as relative standard deviation (RSD) is $\leq 20\%$.

Stability of these CRMs has been listed as 2 year from the introduction date. The materials were processed and are stored at ambient temperature, under nitrogen gas, in 27 -mL glass headspace vials. These materials are expected to be stable for longer than the estimated expiration date. The stability of the powder material will be reevaluated at

time of expiration. If the samples are determined to be stable, the certificates will be extended.

References

AVEKA, Inc.; 2045 Wooddale Drive. Woodbury, MN 55125; Telephone : +1 651 730 1729; <https://www.aveka.com/>

Biosafety Clearing House Living Modified Organism (LMO) Registry
<http://bch.cbd.int/database/lmo-registry/>

Eurofins-GeneScan; 2219 Lakeshore Drive Suite 400, New Orleans, LA 70122; Telephone: +1 504 846 2398; <https://www.eurofinsus.com/food-testing/testing-services/gmo/>

Illinois Crop Improvement Association, 3105 Research Road, Champaign, IL 61826; Telephone: +1 217 359 4053 Fax: +1 217 359 4075; <https://www.ilcrop.com/>

ISO 9001:2015, Quality Management Systems – Requirements

ISO 17025:2005 and ISO 17025:2017, General Requirements for the Competence of Testing and Calibration Laboratories

ISO Guide 17034:2016 (E) General requirements for the competence of reference material producers

Linsinger, T., Pauwels, J., van der Veen, A. *et al.* Homogeneity and stability of reference materials. *Accred Qual Assur* **6**, 20–25 (2001). <https://doi.org/10.1007/s007690000261>

Regulation (EC) No 1829/2003 of the European Parliament and of the Council of 22 September 2003 on genetically modified food and feed (<https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32003R1829&from=en>)

Remund K, Simpson R, Laffont J-L, Wright D, and Gregoire S. Seedcalc8. 2008. (<https://www.seedtest.org/en/statistical-tools-for-seed-testing-content---1--3449--1102.html>)