

Certificate of Analysis			
AOCS 0911-B, non-modified soybean (1g vial)			
Certified Presence	Certified Value	Measurement Uncertainty	Test Method
CV127 soybean absent	0 g/kg	1 g/kg	event-specific real-time PCR
<p>Description: This is the first batch of non-modified soybean CRM prepared by AOCS for BASF Plant Science L.P. A total of 3000 seeds were evaluated and all tested negative for CV127 soybean. The measurement uncertainty is the expanded uncertainty using the value of the upper bound of impurity. The standard uncertainty can be obtained by dividing the expanded uncertainty by $2\sqrt{3}$ (rectangular distribution). This material is for limited purposes only: see "Intended Use" and "Terms and Conditions."</p>			
<p>This certificate is valid through: July 2025</p> <p>This validity date may be extended when further evidence of stability becomes available. Customers will be notified by AOCS if a stability issue arises at AOCS Headquarters.</p> <p>Introduced: 29 February 2012 Revised: 17 February 2013, 24 July 2013, 14 February 2014, 17 February 2015, 31 August 2015, 2 August 2017, 19 July 2018, 19 June 2019, 22 October 2019, 9 July 2020, 1 July 2021, 6 July 2022, 24 July 2023, 8 July 2024, 8 November 2024</p>			
<p>Technical Services Manager Denise Williams</p>		<p>Chief Executive Officer Patrick J. Donnelly</p>	
<p>AOCS Mission: AOCS advances the science and technology of oils, fats, proteins, surfactants and related materials, enriching the lives of people everywhere.</p>			

Characterization

Product Description AOCS 0911-B has been prepared by AOCS from soybean devitalized seed. AOCS 0911-B is available in 10-mL glass headspace vials containing approximately 1g of material. Users are informed that this reference material has been produced from devitalized seed of non-modified soybean delivered by BASF Plant Science L.P. The non-modified soybean used in the preparation of AOCS 0911-B resulted from several cycles of self-pollination of the conventional line.

Homogeneity The homogeneity of AOCS 0911-B is related to the purity of the seeds. 3000 out of 3000 seeds tested negative for the CV127 soybean event by real-time PCR by Eurofins-GeneScan, Metairie, LA (an ISO 17025 accredited laboratory). Based on the sample impurity of 0% as determined using SeedCalc8, the batch was expected to be homogenous.

In addition, a total of ten packaged samples, 1g each, were tested using event-specific qualitative PCR methods developed and validated by BASF Plant Science L.P. Test results received from Eurofins-GeneScan, Metairie, LA (an ISO 17025 accredited laboratory) for qualitative, event-specific analysis were all consistent with the reported absence of CV127 traits.

Stability AOCS Certified Reference Materials are assessed for transport (short-term) and long-term stability and tested for stability on an annual basis and certificates may be extended based on the outcome of this testing. Customers may request extended certificates, but they are informed that results are based on samples that are obtained from AOCS' inventory. AOCS cannot guarantee the integrity of samples outside of our control.

Analytical Method Used for Certification

Eurofins-GeneScan, Metairie, LA (an ISO 17025 accredited laboratory) performed event-specific real-time PCR for non-modified to validate the absence of the CV127 traits. Purity and stability results were used to determine the expanded uncertainty of the certified value.

Warnings and Precautions

This product is for laboratory use only and is not for consumption. The user of this CRM should follow safety requirements and rules issued by voluntary organizations and government agencies expert in the field of laboratory safety.



Intended Use

This CRM, AOCS 0911-B, is intended for use as quality control material or calibrant in methods for the detection, identification, and/or quantification of biotechnology-derived events.

Instructions for Use

Upon receipt the product should be stored in a sealed container in the dark and at ambient or cooler conditions. The product may have settled during shipment, therefore, thoroughly mix the CRM before use to ensure homogeneity.

If the user of this CRM intends to use it multiple times, proper protocols must be followed to ensure that the sample retains its integrity. Use a clean laboratory spatula to remove the intended sample amount. After the sample has been removed, flush the headspace of the vial with nitrogen gas, then replace the rubber stopper. Place a new 20 mm tear-off aluminum unlined seal on top of the rubber stopper and crimp it to the vial by using a crimping tool. Store the CRM in the dark and at ambient or cooler conditions and repeat this process for all subsequent uses.

Sample size:

The recommended **sample intake** suitable for DNA extraction and real time PCR is 1 g.

Protocols have been followed to ensure that this CRM is absent of the traits it has been tested against. Please follow all instructions on this certificate to prevent contamination and be sure to store the CRM under the proper conditions.

Note: The AOCS 0911-B certification report is available online and a paper copy will be supplied upon request.

Terms and Conditions (i– ix)

- i. The CRM AOCS 0911-B shall be used solely 1) in assays for (a) detecting the absence of CV127, or (b) quantification of CV127; or 2) for determining whether an assay cross-reacts with CRM AOCS 0911-B. CRM AOCS 0911-B shall be used for no other purpose. Specifically, the CRM may not be used to develop a detection method for non-modified nor trait(s) present therein. No other rights are conveyed by the sale of the CRM 0911-B to any purchaser, including any rights to any pending or granted BASF Plant Science L.P Patents or other BASF Plant Science L.P Intellectual Property that may protect



the CRM or non-modified or trait(s) present therein or a detection method for non-modified powder.

- ii. Neither the CRM AOCS 0911-B nor any extract or portion thereof shall be resold or redistributed by any purchaser, unless the resale or redistribution is required by national law in force in the purchaser's country.
- iii. Neither the CRM AOCS 0911-B, nor any extract or portion thereof, shall be used for human or animal consumption or human or animal trials.
- iv. Neither the CRM AOCS 0911-B nor non-modified DNA, nor any part of either of these, shall be used for transformation or breeding.
- v. No characterization or derivation of CV127, of the trait or traits present in CRM AOCS 0911-B, or of the CRM AOCS 0911-B shall be performed, except as allowed for in section (i).
- vi. All assay activities undertaken using the CRM AOCS 0911-B shall be conducted in strict compliance with all Applicable Laws governing such activities and shall comply with conditions of all permits and authorizations which may be required for such activities; and such activities shall be strictly limited to assays in contained facilities, for example, laboratories.
- vii. Prior to disposal of any used or excess CRM AOCS 0911-B or residues thereof, such material or residue must be treated in a manner that degrades the CRM material, such as by autoclaving.
- viii. CRM AOCS 0911-B shall not be exported nor re-exported in violation of any Applicable Laws or without securing any necessary export or import clearances or permits.
- ix. THE CRM 0911-B IS PROVIDED FOR THE PURPOSE OF IDENTIFYING, DETECTING AND QUANTIFYING CV127 AND FOR NO OTHER PURPOSE. AOCS HAS TESTED CRM 0911-B WITHIN THE PAST 12 MONTHS AND STATES THAT IT HAS DETERMINED IT TO BE OF SUFFICIENT QUALITY AND FIT FOR THE PURPOSES STATED HEREIN. NO BASF Plant Science L.P WARRANTY IS PROVIDED WHETHER EXPRESS OR IMPLIED, IN RELATION TO THE CRM AOCS 0911-B AND BASF Plant Science L.P MATERIALS. BASF Plant Science L.P MAKES NO REPRESENTATION OR WARRANTY THAT THE USE OF THE CRM AOCS 0911-B, WHETHER BEFORE, OR AFTER THE EFFECTIVE DATE OF THE APPLICABLE CERTIFICATE, WILL NOT INFRINGE ANY PATENT OR OTHER PROPERTY RIGHTS.

