

## Certificate of Analysis

### AOCS 0610-A7, FG72 soybean seed genomic DNA



Certified Presence	Certified Value	Measurement Uncertainty	Test Method
Homozygous FG72 DNA present	>992.5 ng/μg FG72 DNA	3.6 ng/μg	event-specific real-time PCR

**Description:** This is the seventh batch of FG72 soybean CRM prepared by AOCS for Syngenta Crop Protection, LLC ("Syngenta"). Syngenta is acting as a contractor on behalf of the owner of the FG72 event M.S. Technologies, LLC ("MS Tech"). The certified value is based on a sample purity of 100% (396 out of 396 seeds tested positive for FG72). With 95% confidence, the true value is > 992.5 ng/μg. The measurement uncertainty is the expanded uncertainty with a coverage factor of 1.65 and confidence level of 95%. This material is for limited purposes only: see "Intended Use" and "Terms and Conditions."

### This certificate is valid through: March 2026

This validity may be extended when further evidence of stability becomes available. Customers will be notified by AOCS if a stability issue arises at AOCS Headquarters.

**Introduced:** 9 July 2024 **Revised:** 7 March 2025

	
Technical Services Manager Denise Williams	Chief Executive Officer Patrick J. Donnelly

AOCS Mission: AOCS advances the science and technology of oils, fats, proteins, surfactants and related materials, enriching the lives of people everywhere.

## Characterization

**Product Description** AOCS 0610-A7 has been prepared by AOCS from homozygous genomic soybean DNA extracted from clean seeds. AOCS 0610-A7 is available in sterile, 0.5 mL skirted screw-cap self-sealing tubes containing approximately 10 µg of DNA. Each sample contains approximately 18.7 µL at a concentration of 537 ng/µL. Users are informed that this reference material has been produced from genomic DNA extracted from clean seeds of FG72 soybean by Syngenta. The FG72 soybean used in the preparation of AOCS 0610-A7 resulted from several cycles of self-pollination of the FG72 line. Seed for homozygous FG72 soybean were provided to Syngenta by MS Tech.

**Homogeneity** Ten packaged samples were tested using event-specific qualitative PCR methods developed and validated by Bayer CropScience (the original owner of the Event FG72). Test results received from FoodChain ID, Chantilly, VA (an ISO 17025 accredited laboratory) for qualitative, event-specific analysis were all consistent with the reported presence of the FG72 trait.

**Stability** AOCS Certified Reference Materials are tested for stability on an annual basis and certificates may be extended based on the outcome of this testing. Customers may request extended certificates, but they are informed that results are based on samples that are obtained from AOCS' inventory. AOCS cannot guarantee the integrity of samples from batches that we no longer have inventoried.

## Analytical Method Used for Certification

FoodChain ID, Chantilly, VA (an ISO 17025 accredited laboratory) performed event-specific real-time qualitative PCR to validate the presence of the FG72 trait.

## Warnings and Precautions

This product is for laboratory use only and is not for consumption. The user of this CRM should follow safety requirements and rules issued by voluntary organizations and government agencies expert in the field of laboratory safety.

## Intended Use

This CRM, AOCS 0610-A7, is intended for use as quality control material or calibrant in methods for the detection, identification, and/or quantification of biotechnology-derived events.



## Instructions for Use

Upon receipt the product should be stored in a sealed container at +4 °C in the dark. The product may have redistributed during shipment, therefore, centrifuge the vial to aggregate the sample at the bottom of the vial before use.

If the user of this CRM intends to use it multiple times, proper protocols must be followed to ensure that the sample retains its integrity. Use a clean pipette to extract the intended sample amount. Close the lid on the vial and rewrap with new parafilm. Store the CRM +4 °C in the dark and repeat this process for all subsequent uses.

Recommended test portion: minimum of 200 ng or the test portion recommended in your validated method

Protocols have been followed to ensure that this CRM has the presence of the FG72 trait. Please follow all instructions on this certificate to prevent contamination of the CRM.

Note: The AOCS 0610-A7 certification report is available online and a paper copy will be supplied upon request.

## Terms and Conditions (i– ix)

- i. The CRM AOCS 0610-A7 shall be used solely 1) in assays for (a) detecting the presence of FG72, or (b) quantification of FG72; or 2) for determining whether an assay cross-reacts with CRM AOCS 0610-A7. CRM AOCS 0610-A7 shall be used for no other purpose. Specifically, the CRM may not be used to develop a detection method for FG72 nor trait(s) present therein. No other rights are conveyed by the sale of the CRM 0610-A7 to any purchaser, including any rights to any pending or granted Syngenta or MS Tech Patents or other Syngenta or MS Tech Intellectual Property that may protect the CRM or FG72 or trait(s) present therein or a detection method for FG72.
- ii. Neither the CRM AOCS 0610-A7 nor any extract or portion thereof shall be resold or redistributed by any purchaser, unless the resale or redistribution is required by national law in force in the purchaser's country.
- iii. Neither the CRM AOCS 0610-A7, nor any extract or portion thereof, shall be used for human or animal consumption or human or animal trials.
- iv. Neither the CRM AOCS 0610-A7 nor FG72 DNA, nor any part of either of these, shall be used for transformation or breeding.
- v. No characterization or derivation of FG72, of the trait or traits present in CRM AOCS 0610-A7 shall be performed, except as allowed for in section (i).



- vi. All assay activities undertaken using the CRM AOCS 0610-A7 shall be conducted in strict compliance with all Applicable Laws governing such activities and shall comply with conditions of all permits and authorizations which may be required for such activities; and such activities shall be strictly limited to assays in contained facilities, for example, laboratories.
- vii. Prior to disposal of any used or excess CRM AOCS 0610-A7 or residues thereof, such material or residue must be treated in a manner that degrades FG72 in the CRM material, such as by autoclaving.
- viii. CRM AOCS 0610-A7 shall not be exported nor re-exported in violation of any Applicable Laws or without securing any necessary export or import clearances or permits.
- ix. The CRM AOCS 0610-A7 is provided for identifying, detecting and quantifying the event FG72 and for no other purpose. AOCS has tested CRM AOCS 0610-A7 within the past 12 months and states that it has determined it to be of sufficient quality and fit for the purposes stated herein. No Syngenta or MS Tech warranty is provided, whether express or implied, in relation to the CRM AOCS 0610-A7 and Syngenta or MS Tech materials. Syngenta and MS Tech make no representation or warranty that the use of the CRM AOCS 0610-A7, whether before, or after the effective date of the applicable certificate, will not infringe any patent or other proprietary rights.

