AOCS
STREET ADDRESS
3356 Big Pine Trail, Ste. C&D,
Champaign, IL 61822 USA
MAILING ADDRESS
PO Box 7230,
Champaign, IL 61826 USA
+1 217-359-2344
general@aocs.org
aocs.org

Certified Reference Materials

AOCS 0406-D2

Report of the certification process for

MON88017

Maize

Second Batch

OECD Unique ID MON-88Ø17-3

Zoe Serelis
Technical Services Manager

Tiffanie West
Technical Services Director

A2LA ACCREDITED
ISO 17034:2016
A2LA Certificate 3438.01

Legal Notice

Neither AOCS nor any person acting on behalf of AOCS is responsible for the use which might be made of the following information.

AOCS Mission Statement

AOCS advances the science and technology of oils, fats, proteins, surfactants, and related materials, enriching the lives of people everywhere.

More information regarding AOCS is available at www.aocs.org

©AOCS, 2026

Contents

Abstract 4
Acknowledgements 5
Glossary 6
Introduction 7
Material Processing 7
Trait Verification to Certify Presence of MON 88017 7
Certified Value and Measurement Uncertainty 8
Homogeneity 9
Stability 10
References 13

©AOCS, 2026

Abstract

This report describes the preparation and certification of the maize CRM AOCS 0406-D2 produced by AOCS Technical Services in 2019. The CRMs have been prepared according to ISO 17034:2016 and are intended to serve as control material for third party testing of maize for transformation events. The maize MON 88017 powder was provided by Bayer CropScience, St. Louis, MO. It was prepared by grinding the bulk seed at Bayer CropScience. The certified value of AOCS 0406-D2 was based on the purity of the bulk seed material and with 95% confidence, the true value is ≥ 928 g/kg. The powder was aliquoted and bottled in 27-mL glass headspace vials and sealed under a nitrogen gas environment at Illinois Crop Improvement Association. The presence of MON 88017 in AOCS 0406-D2 was verified using event-specific, qualitative PCR analysis by Eurofins-GeneScan, New Orleans, LA (an ISO 17025-accredited laboratory). Homogeneity was verified on random vials of AOCS 0406-D2 using digital PCR analysis by Bayer CropScience. CRM samples should be stored in a dry, sealed container at ambient or cooler conditions in the dark.

©AOCS, 2026

Acknowledgements

The authors would like to express sincere appreciation and gratitude to several individuals and their companies for support and guidance throughout this project. Thanks go to Jack Milligan, Bayer CropScience, for offering AOCS the opportunity to manufacture and distribute these products; to Sandra Harrison and Charlie Drennan at Illinois Crop Improvement Association for packaging the samples; and to Frank Spiegelhalter, Greg Ditta, E. Pearce Smith, and Daniel Thompson, Eurofins-GeneScan for event-specific, qualitative PCR analysis including the provision of information on running the analyses and interpreting the results.

©AOCS, 2026

Glossary

AOCS	American Oil Chemists' Society
Conventional Crop	Crop variety with no history of modern biotechnology and is produced through plant-breeding techniques that rely on selecting and mating parent plants possessing promising traits and repeatedly selecting for superior performance among their offspring
DNA	Deoxyribonucleic Acid
Detection Limit	Lowest level at which target DNA can be detected in a sample and be reliably tested by PCR methods. It is typically expressed as a percentage: the ratio of the number of modern biotechnology derived genomes to the number of crop genomes times 100 percent
EC	European Commission
Genome	The full set of genes and associated DNA characteristic of an organism
GMO	Organism that has had genetic sequences modified using molecular-level techniques
ISO	International Organization for Standardization
PCR	Polymerase Chain Reaction: technique used to determine whether a sample of plant tissue contains a particular DNA sequence. PCR relies on primer sets that zero in on a particular target DNA sequence and a special DNA-copying enzyme (DNA polymerase) that makes enough copies of the target sequence for identification and measurement
Qualitative PCR	PCR methods that determine the presence or absence of a specific target DNA sequence at a particular level of detection
Quantitation Limit	Lowest level at which the amount of target DNA sequence in a sample can be reliably quantified. It is typically expressed as the ration of the number of transgenic genomes to the number of crop genomes times 100 percent
Quantitative PCR	PCR methods that estimate the relative amount of target DNA sequence in a mixture of DNA molecules
Trait: MON88017	Coleopteran insect resistance and glyphosate herbicide tolerance

©AOCS, 2026

Introduction

Plant genetic modification is an extension of traditional plant breeding. It allows plant breeders to develop crops with specific traits including insect, disease, and herbicide resistance; processing advantages; and nutritional enhancement. An important component for identifying these new traits is a Certified Reference Material (CRM) created from leaf, seed, or grain containing the new trait as well as a CRM created from the conventionally bred matrix. The European Commission has mandated that from 18 April 2004, a method for detecting a new event derived from transgenic technology and Certified Reference Material must be available before the EC will consider authorizing acceptance of a new crop derived from transgenic technology. Several nations outside Europe also require grain and ingredients to be labeled above a threshold level before accepting a shipment.

To meet the above regulatory requirements for GMO determination, AOCS 0406-D2 was manufactured from maize according to ISO 17034:2016 and in accordance with EC No 1829/2003, EC No 641/2004 and EC No 619/2011. This CRM is available from AOCS.

Material Processing

MON 88017 maize seeds used to prepare AOCS 0406-D2 were hemizygous through successive breeding generations, and the donor for the MON 88017 maize event was the female parent. Bayer CropScience milled ~10 kg of MON 88017 maize seed. All of the seed powder was passed through a 850 µM mesh sieve. The seed powder was delivered to AOCS who contracted Illinois Crop Improvement Association for packaging the samples. The powder was aliquoted and bottled in 27-mL glass headspace vials and sealed under a nitrogen gas environment.

Trait Verification to Certify Presence of MON 88017

The presence of the MON 88017 trait was assessed on 10 random vials of AOCS 0406-D2. AOCS used the Random Number Generator function of Microsoft Excel to select samples for verification of trait presence. Sample numbers that were randomly selected were sent to Eurofins-GeneScan, New Orleans, LA (an ISO 17025-accredited

©AOCS, 2026

laboratory) for event-specific, qualitative PCR analysis to verify the presence of MON 88017 in the samples (Table 1).

Table 1. Trait verification testing on AOCS 0406-D2 MON 88017 maize performed by Eurofins-GeneScan, New Orleans, LA (an ISO 17025-accredited laboratory).

AOCS 0406-D2 Sample	Trait MON 88017 Presence
Sample #0758	Positive
Sample #0302	Positive
Sample #0244	Positive
Sample #0180	Positive
Sample #0431	Positive
Sample #0641	Positive
Sample #0058	Positive
Sample #0546	Positive
Sample #0576	Positive
Sample #0385	Positive

Certified Value and Measurement Uncertainty

The genetic purity of the seed lot used to produce AOCS 0406-D2 was assessed by Bayer CropScience. A total of 319 maize seeds were subjected to individual seed testing for the presence of MON 88017 by qualitative event-specific PCR. 319 of the 319 seeds tested positive for the presence of MON 88017.

Purity estimation was calculated using SeedCalc8 (Remund *et al*., 2008) and corresponds to the lower bound of true % purity. The % purity in the sample was 100% when 319 seeds were tested. Using a 95% confidence level, the true % purity of the MON 88017 seed lot was at least 99.1%. Consequently, with 95% confidence, the true value is ≥ 928 g/kg.

The measurement uncertainty (U_{CRM}) is the expanded uncertainty with a coverage factor of 2 and a confidence level of 95%. It is obtained by combining the uncertainties from the purity assessment ($u_{char,rel}$), the homogeneity assessment ($u_{bb,rel}$), the transport stability assessment ($u_{sts,rel}$) and the long-term stability assessment ($u_{lts,rel}$):

©AOCS, 2026

$$u_{CRM,rel} = \sqrt{u_{char,rel}{}^2 + u_{bb,rel}{}^2 + u_{sts,rel}{}^2 + u_{lts,rel}{}^2}$$

$$U_{CRM} = 2 \times u_{CRM,rel} \times 1000\ g/kg$$

Consequently, the expanded measurement uncertainty for AOCS 0406-D2 is -72 g/kg.

Homogeneity

The homogeneity of AOCS 0406-D2 is related to the purity of the seeds. 319 out of 319 seeds tested positive for the MON 88017 maize event by event-specific PCR. Based on the sample purity of 100%, as determined using SeedCalc8, the batch was expected to be homogenous.

To further confirm homogeneity, ten vials of AOCS 0406-D2 (randomly selected as described above) were provided by AOCS to Bayer CropScience. Homogeneity was assessed using the MON 88017 specific quantitative PCR method (gmo-crl.jrc.ec.europa.eu/method-validation/details/all/41/MON%2088017) that was adapted for digital PCR (dPCR), which has the advantage over qPCR of quantifying targets without the need for calibration curves. For each of the 10 CRM vials analyzed, there were 2 independent DNA extractions. Each DNA extraction was subject to 3 dPCR replicates. The data produced from these dPCR reactions provided the numeric copies of MON 88017 and the numeric copies of *hmg*, a maize specific endogenous reference gene. The property value assessed here is defined as the ratio between copies of the MON 88017 target and copies of the *hmg* target.

The digital PCR data was used to evaluate the within-unit and between-unit homogeneity of AOCS 0406-D2 to ensure that the property value is valid within vials of CRM and between vials of CRM. The CRM will be determined to be homogeneous if the within-unit relative standard deviation ($\mathrm{RSD_w}$) and between-unit relative standard deviation ($\mathrm{RSD_b}$) are both ≤20%.

Quantification of between-unit (vial/sample) inhomogeneity was undertaken by analysis of variance (ANOVA), which separates the between-unit variation from the within-unit variation. Preliminary analysis showed that there is no significant variation between the

©AOCS, 2026

two DNA extractions within each vial, so the DNA extraction effect was not considered in the analysis. That is, all replicates for each vial were treated as independent observations regardless of which DNA extraction they were from.

Within-unit relative standard deviation ($\mathrm{RSD_w}$), between-unit relative standard deviation ($\mathrm{RSD_b}$) were calculated as:

Within-unit RSD: $RSD_w = \frac{\sqrt{MS_{within}}}{\bar{y}}$

Between-unit RSD: $RSD_b = \frac{\sqrt{\frac{MS_{between} - MS_{within}}{n}}}{\bar{y}}$

where,

$\mathrm{MS_{within}}$ within-unit mean square from an ANOVA
$\mathrm{MS_{between}}$ between-unit mean square from an ANOVA
$\bar{y}$ mean of all results of the homogeneity study
n mean number of replicates per unit (6 for MON 88017)

Table 2. The within-unit relative standard deviation (RSD_w), and the between-unit relative standard deviation (RSD_b) for vials of AOCS 0406-D2.

CRM	RSD_w [%]	RSD_b [%]
AOCS 0406-D2	6.2	3.4

This confirms the homogeneity of AOCS 0406-D2.

Stability

Time, temperature and light are regarded as the most relevant influences on the stability of CRM (Linsinger, et al., 2001). The influence of light is mitigated by shipping and storing the vials in boxes, thus minimizing the possibility of degradation due to light. The influence of temperature is mitigated by storing the vials in a temperature-controlled room, and shipping vials at ambient temperature.

The effect of temperature and time are investigated.

A transport (short-term) stability study is conducted to assess the stability of maize CRM during transport. The temperature and time conditions in the study cover the typical

©AOCS, 2026

conditions and the not so rare situations. The outcome of the study is considered transferable to other CRMs of similar property. Samples were subject to 3 different temperatures (4 ˚C (fridge), 25 ˚C (ambient), 60 ˚C (oven)) for 4 different durations (0, 1, 2, and 4 weeks). The study concluded that samples are stable at 4 ˚C (fridge) and 25 ˚C (ambient) for 4 weeks. The estimated uncertainty contribution from transport (short-term) stability is 1.0%.

A long-term stability study is conducted to assess the stability of maize CRM during storage. Samples are stored at 25 ˚C (ambient) and the stability of the sample is monitored as long as the samples is available. The storage temperature study is 25 ˚C and the length of time to be studied is 10 years. The outcome of the study is considered transferable to other CRMs of similar property. In the initial 1-year stability study, samples were subject the storage condition for 4 different durations (0, 1, 3, 6 and 12 months). The study concluded that samples are stable at 25 ˚C (ambient) for 12 months. The estimated uncertainty contribution from long-term stability is 0.42%.

CRM stability over time will be analyzed by repeating the homogeneity study described above at a chosen shelf life of approximately every 24 months. The 24-month shelf life of CRM is chosen because the influence of analytical variation can be reduced by increasing the length of the stability study (Linsinger, et al., 2001).

The initial ratio between the number of copies of the GM event and the number of copies of the endogenous reference gene from the homogeneity study will establish the base line for the stability study. The ratio at each 24-month interval will be compared to the ratio established in the homogeneity study. The CRM will be determined to be stable if the variability of the ratios, determined as relative standard deviation (RSD) is ≤ 20%.

Stability of these CRMs has been listed as 2 years from the certification date. The materials were processed and are stored at ambient temperature, under nitrogen gas, in 27 -mL glass headspace vials. These materials are expected to be stable for longer than the estimated expiration date. The stability of the powder material will be reevaluated at time of expiration. If the samples are determined to be stable, the certificates will be extended.

©AOCS, 2026

©AOCS, 2026

References

Eurofins-GeneScan; 2219 Lakeshore Drive, Suite 400, New Orleans, LA 70122; Telephone: +1 504 297 4330 Toll Free: +1 866 535 2730 Fax: +1 504 297 4335 https://www.eurofinsus.com/food-testing/testing-services/gmo/

Illinois Crop Improvement Association, 3105 Research Road, Champaign, IL 61826; Telephone: +1 217 359 4053 Fax: +1 217 359 4075; www.ilcrop.com/

ISO 17034:2016 (E) General requirements for the competence of reference material producers

ISO 17025:2005 and ISO 17025:2017, General Requirements for the Competence of Testing and Calibration Laboratories

International Seed Testing Association, International Rules of Seed Testing: Seed Science and Technology Rules, 2012

Regulation (EC) No 1829/2003 of the European Parliament and of the Council of 22 September 2003 on genetically modified food and feed; https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX%3A32003R1829&from=en

Remund K, Simpson R, Laffont J-L, Wright D, and Gregoire S. Seedcalc8. 2008 https://www.seedtest.org/en/services-header/tools/statistics-committee/statistical-tools-seed-testing.html

Linsinger, et. al.," Homogeneity and Stability of reference materials", Accreditation and Quality Assurance, 2001

©AOCS, 2026